



2018 Rapid DNA Maturity Assessment

Erica Romsos & Peter Vallone


Applied Genetics Group, NIST

November 7, 2018



Steps in Forensic DNA Analysis

- Steps Involved
- Collection
 - Extraction**
 - Quantitation**
 - Multiplex PCR
 - STR Typing
 - Interpretation of Results



1-2 day process (a minimum of ~8 hours) with current laboratory procedures and technology



NIST: Rapid DNA Maturity Assessments

- Collection and distribution of samples to all participating laboratories
 - 2013: 3 labs, 150 samples
 - 2014: 7 labs, 280 samples
 - 2018: 9 participants, 240 samples
- Coordination of all testing sites to include return of all data to NIST for analysis and review
 - Includes IRB approval for swab collection and MTA for distribution
- Analysis and compilation of all data
- Summary of results presented across multiple meetings within the forensic and biometric communities



FORENSICS @ NIST

#NISTForensics

2013 Rapid DNA Maturity Assessment

- Summer 2013 we felt confident that we could carry out an interlab assessment of the R-DNA prototypes
- Data was collected and analyzed in August 2013
 - All instruments had the same version of software and scripts for testing
- Three federal testing sites (7 individual instruments)
 - 50 single source reference buccal swabs were provided by NIST for each instrument tested
- 350 single source reference buccal swabs tested
 - Success defined as the automated calling of the 13 CODIS core STR loci
- **Overall success = 88.3%**




FORENSICS @ NIST

#NISTForensics

2014 Rapid DNA Maturity Assessment

- Fall 2014 assessment of the current status of rapid DNA typing technology for the CODIS Core Loci
 - In support of lab and future booking station Rapid DNA implementation
- Many modifications to both hardware and software were made between 2013 and 2014
- 7 participating laboratories (11 independent instruments)
 - 20 single source reference buccal swabs tested
- 280 single source reference buccal swabs tested
 - Success defined as automated (lights out) calling of the 13 CODIS core loci
- **Overall Success = 76.1%**



FORENSICS @ NIST

#NISTForensics

2018 Rapid DNA Maturity Assessment


- **Goal:** To measure the status of rapid DNA typing technology for the 20 CODIS core loci in support of booking station Rapid DNA implementation
- Rapid DNA instruments capable of genotyping the 20 CODIS core loci were eligible for participation
- 20 single source reference buccal swabs were distributed to participating laboratories



FORENSICS @ NIST

#NISTForensics

2018 Rapid DNA Maturity Assessment Study Outline



Timeline of 2018 Maturity Assessment

February 2018-Present: Buccal samples collected at NIST and stored at RT (>900 swabs have been collected from >50 unique individuals)

May 2018: Official call for Participants in the 2018 Rapid DNA Maturity Assessment for the CODIS core 20 loci

July 2018: Samples shipped to participants

August 2018: Data returned to NIST for analysis



20 Swabs provided

Participants	Instrument Platforms	Chemistry	Independent Instruments	Total Samples Tested	Analysis Method
Federal	ANDE 6C 	FlexPlex	5	100	Rapid DNA Analysis
State	IntegenX RapidHIT 200 	GlobalFiler Express	3	60	Modified Rapid DNA Analysis
Police	IntegenX RapidHIT ID 	GlobalFiler Express	4	80	Modified Rapid DNA Analysis
9 Participants	3 Platforms		12 Instruments	240 Samples	

Success Metrics

- Success was measured by **complete and concordant genotypes** produced for the 20 CODIS core loci
- Allele calls by the integrated rapid DNA devices were compared to lab generated profiles for concordance
 - Fusion 6C, PP21, GFE on a 3500xL
- Two interpretation parameters were implemented
 - Rapid DNA Analysis: **Without human intervention**
 - Modified Rapid DNA Analysis: **Expert interpretation and analysis of electropherogram** <https://www.fbi.gov/file-repository/addendum-to-qas-for-rapid-dna.pdf/view>



FORENSICS @ NIST

#NISTForensics

Rapid DNA Analysis-ANDE

Automated (lights-out) analysis without human intervention

Effective June 1, 2018, the following Rapid DNA system is approved for use at NDIS by an accredited forensic DNA laboratory:

Rapid DNA Analysis System for Accredited Laboratory Use

Component	Name	Part/Version Number
Rapid DNA Instrument	ANDE 6C Instrument	A0120001003
Typing Kit	FlexPlex27	FlexPlex27
Cartridge	ANDE A-Chip (FlexPlex)	A0210001057
System Software	ANDE System Software	2.0.6
Expert System Software	ANDE Expert System	2.0.5

To date, ANDE 6C is the only rapid DNA system to be NDIS approved for automated rapid DNA analysis

<https://www.fbi.gov/services/laboratory/biometric-analysis/codis/rapid-dna>




FORENSICS @ NIST

#NISTForensics

CODIS 20 Success: Automated Analysis

n=85



Average Success: 85%

Instruments



FORENSICS @ NIST

#NISTForensics


Modified Analysis Parameters

- Samples were manually reviewed (n=240)
 - ANDE profiles were analyzed in GeneMapper IDX v1.5
 - RapidHIT profiles were reviewed in GeneMarker HID v2.8.2
 - Heterozygote balance filter set at 0.25
- After manually interpreting a profile (PHR >0.25, low AT, recovery of “lost” data), concordance was checked against the laboratory generated reference profile
- Success was determined by **complete and concordant** profiles for the 20 CODIS core loci



FORENSICS @ NIST

#NISTForensics




FORENSICS @ NIST

#NISTForensics

Unrecovered Samples

n=23



FORENSICS @ NIST

#NISTForensics

Unrecovered Samples

n=23

Instrument Related

Unknown
n=9

Data
Transfer
Failure
n=2



Name	Date modified	Type
DannoGUIState.xml	8/21/2018 2:51 PM	XML Document
GelFill.HV_(14.23).png	8/21/2018 2:51 PM	PNG File
Storyboard_(14.23).txt	8/21/2018 2:51 PM	Text Document
SyringePump_(14.23).csv	8/21/2018 2:40 PM	Microsoft Excel C...

No data was transferred from the instrument to analyze




FORENSICS @ NIST

#NISTForensics



Summary of Success



- **Automated success rate of 85% (n=100)**

- ANDE 6C is the only NDIS approved instrument for automated analysis
- xml files for the unsuccessful profiles were not generated by the instrument (15%)


- **Modified analysis success rate of 90%**

- Manual interpretation of all samples (n=240 samples)
- Increase of success for manually reviewed profiles




FORENSICS@NIST

#NISTForensics



n=217 successful profiles



Additional Metrics Analyzed

Base pair sizing precision

Heterozygote Balance




FORENSICS@NIST

#NISTForensics


Precision-Base Pair Sizing

n=19 ladders for ANDE
n=8 ladders for RH 200

ANDE Size Pooled Standard Deviation



RapidHIT 200 Size Pooled Standard Deviation




FORENSICS @ NIST

#NISTForensics

Heterozygote Balance

ANDE



0.52


95% of the samples were greater than 52%



FORENSICS @ NIST

#NISTForensics

Heterozygote Balance




95% of the samples
were greater than 63%



FORENSICS @ NIST

#NISTForensics

Heterozygote Balance




95% of the samples
were greater than 52%



FORENSICS @ NIST

#NISTForensics



Maturity Assessment Summary

- 12 instruments tested across 9 laboratories
- Total of 240 samples examined
 - 85% success rate for the CODIS 20 using **Rapid DNA Analysis**
 - 90% success rate for the CODIS 20 using **Modified Rapid DNA Analysis**
 - Success ranged from 60% to 100%
 - Precision was below 0.17 bp on for both ANDE 6C and RapidHIT 200
 - Combined heterozygote balance (all three instruments) was above 59%

Acknowledgements

Thank you to our participants

- ANDE
- Arizona Department of Public Safety
- Bensalem Police Department
- Federal Bureau of Investigation Laboratory
- Louisiana State Police Crime Laboratory
- Miami Beach Police Department
- Miami Dade Police Department
- NIST (DHS instruments, run at SNA Intl.)
- U.S. Army Criminal Investigation Laboratory

NIST – Applied Genetics Group

Peter Vallone

Steven Lund

Funding

FBI Biometrics Center of Excellence: Forensic
DNA Typing as a Biometric tool.

Contact Information

erica.romsos@nist.gov



FORENSICS @ NIST

#NISTForensics