

# NIST Validation Studies on the 3500 Genetic Analyzer



**Erica L.R. Butts**

National Institute of Standards and Technology

Innsbruck, Austria

September 5<sup>th</sup>, 2011

# Outline

- Details of the ABI 3500 Genetic Analyzer
- Validation design and results with Identifiler and Identifiler Plus
  - Injection parameters and reaction setup
  - Precision and size standard comparison
  - Concordance and mixture evaluation
- Methodology of setting analytical and stochastic thresholds
- Brief overview of signal normalization

# Details of the ABI 3500

**No lower pump block  
(Fewer air bubbles)**



**Improved sealing for better  
temperature control**



# Primary Differences

	<b>31xx Platforms</b>	<b>3500 Platforms</b>
<b>Laser</b>	Argon ion (AR+) with 488/514 nm wavelength	Single-line 505 nm, solid-state, long-life laser
<b>Power Requirement</b>	220V	110V
<b>File Generated</b>	.fsa files	.hid files
<b>Normalization</b>	None	Instrument-to-instrument; only with AB kits
<b>Optimal Signal Intensity</b>	1500-3000 RFU	4x greater than 31xx platforms

# What is Validation?

**Section 1.1 (SWGDM Revised Validation Guidelines)** Validation is the process by which the scientific community acquires the necessary information to:

- (a) Assess the ability of a procedure to obtain reliable results.
- (b) Determine the conditions under which such results can be obtained.
- (c) Define the limitations of the procedure.

The validation process identifies aspects of a procedure that are critical and must be carefully controlled and monitored.

## Reliability, Reproducibility, Robustness

# Experimental Summary

	Test	Types of Samples Used	Number Examined
Reliability	Size Standard Comparison	16 Allelic Ladders per size standard (LIZ 500 vs. LIZ 600 v2.0)	32
	Injection Parameters	3 samples heterozygous at 15 loci plus Amelogenin 1 ng DNA input	15 3 samples per injection
Reproducibility	Precision	Allelic Ladders	24
		3 samples heterozygous at all 15 loci plus Amelogenin	6
	Concordance	50 genomic DNA samples	60
SRM 2391b: 10 genomic DNA samples			
Robustness	Sensitivity	Dilution series of 3 samples heterozygous at 15 loci plus Amelogenin	84 4 replicates of each dilution series
	Mixtures	Mixture dilution series of 2 samples heterozygous at 15 loci plus Amelogenin	28
		<b>Total Number of Samples</b>	<b>249</b>

Identical experiments for **Identifiler** and **Identifiler Plus**

# Size Standard Comparison

	1	2
A	LIZ 500	LIZ 600 v2.0
B	LIZ 600 v2.0	LIZ 500
C	LIZ 500	LIZ 600 v2.0
D	LIZ 600 v2.0	LIZ 500
E	LIZ 500	LIZ 600 v2.0
F	LIZ 600 v2.0	LIZ 500
G	LIZ 500	LIZ 600 v2.0
H	LIZ 600 v2.0	LIZ 500

Individual master mixes created for LIZ 500 and LIZ 600 v2.0 with **Identifiler/Identifiler Plus** allelic ladders


Injected twice on 3130xl/  
 – Standard injection of 3 kV for 10 seconds

Injected 3 times on 3500  
 – Default Injection of 1.2 kV for 15 seconds


It is important to determine if one size standard can be used consistently on both the 3130xl and 3500 for proper comparison

# Size Standard Comparison

LIZ 500





LIZ 600 v2.0





# Size Standard Comparison



LIZ 600 v2.0 generated the most linear results on both the 3130x/ and 3500 and was used as the size standard on both instruments for remaining testing

# Injection Parameters

- Injection voltage/time:

- 1.2 kV for 15 sec

- 1.2 kV for 10 sec

- 1.2 kV for 7 sec

➔ Identifiler


- 1.2 kV for 5 sec

➔ Identifiler Plus

- 1.2 kV for 3 sec

Standard injection parameters set based on samples with:

1. No pull-up present
2. No drop out present



# Sizing Precision

Identifiler

Identifiler Plus

	1	2	3	4
A	Identifiler	EB	Identifiler Plus	EB
B	Neg	Identifiler	Neg	Identifiler Plus
C	Identifiler	EB	Identifiler Plus	EB
D	Neg	Identifiler	Neg	Identifiler Plus
E	Identifiler	EB	Identifiler Plus	EB
F	Neg	Identifiler	Neg	Identifiler Plus
G	Identifiler	Sample	Identifiler Plus	Sample
H	Sample	Identifiler	Sample	Identifiler Plus

Identifiler and Identifiler Plus allelic ladders in checkerboard pattern

Neg: PCR blank

– PCR primers + water



EB: Extraction blank

– PCR primers + extraction eluent

Sample: 1 ng heterozygous sample at 15 loci plus Amelogenin

Injected 3 times with the newly determined injection parameters

# Precision of Base Pair Sizing




**No significant difference between 3130x/ and 3500**

**No significant difference between Identifiler and Identifiler Plus**

# Concordance and Mixtures

- 60 samples concordant between 3130x/ and 3500
  - Total of 1689 alleles examined
- Minor component identified correctly in a 1:10 mixture ratio



# Different Threshold Overview

**Example values**  
(empirically determined  
based on own internal  
validation)

**Called Peak**  
(Greater confidence a sister  
allele has not dropped out)

350 RFUs

**Called Peak**  
(Cannot be confident  
dropout of a sister allele  
did not occur)

**Stochastic Threshold**


The value above which it is  
reasonable to assume that  
allelic dropout of a sister  
allele has not occurred

150 RFUs

**Analytical Threshold**

Minimum threshold for data  
comparison and peak  
detection in the DNA typing  
process

Peak not  
considered  
reliable




**Noise**


# Analytical Threshold Methodology

- Baseline noise values calculated with data from the sensitivity study (DNA dilution series)
  - Threshold set at 1 RFU for all dye channels
  - Remove calls for all alleles and artifacts (stutter, n+4, pull-up, etc.)
- 4 methods for evaluation of analytical thresholds calculated
- Analytical Threshold: Average RFU + (10 x Standard Deviation)

# Different Thresholds



Single thresholds for all dye channels assumes all dye channels have the **same** amount of noise



Can cause data to fall below the analytical threshold and not be called


Dye-specific thresholds take into consideration that all dye channels do not have the same level of noise

Can increase the amount of data that is callable



# One Threshold vs. Dye Specific Thresholds

- Evaluation of data to determine the statistical difference between dye channel analytical thresholds
- Calculated statistical difference using a z-test
- If negative: **Not** statistically different
  - Error bars overlap
  - One standard analytical threshold can be applied to all dyes
- If positive: Statistically different
  - Error bars do not overlap
  - Dye specific analytical thresholds need to be applied



n=84 samples

# Analytical Threshold Calculation

## Identifiler

Dye Channel	Average RFU	Stdev	Min RFU	Max RFU	Calculated Noise (RFU)
Blue	9	8.4	1	66	93
Green	13	11.5	3	84	128
Yellow	22	11.6	4	88	138
Red	28	8.8	10	80	116

Single Threshold:  
140 RFU

Dye-Specific:  
Rounded to  
nearest 5 RFU

## Identifiler Plus

Dye Channel	Average RFU	Stdev	Min RFU	Max RFU	Calculated Noise (RFU)
Blue	10	4.6	3	68	55
Green	16	5.6	3	78	72
Yellow	24	7.9	7	63	103
Red	31	8.9	7	81	120


Single Threshold:  
120 RFU

Dye-Specific:  
Rounded to  
nearest 5 RFU

- Statistical difference was calculated between dye channels using a z-test
- Statistically each dye channel is different for both **Identifiler** and **Identifiler Plus**
  - Must be treated independently

# Threshold Comparison

Total of 560 alleles examined (50 pg, 30 pg, and 10 pg) where dropout was observed



**14.8% of the total possible allele calls were lost** using a single threshold rather than using dye-specific thresholds with **Identifiler**

**22.0% of the total possible allele calls were lost** using a single threshold rather than using dye-specific thresholds with **Identifiler Plus**


# Setting Stochastic Methodology

- Calculated with data from the sensitivity study (DNA dilution series) analyzed with dye specific analytical thresholds
- Examination of sample amounts where dropout is observed (50 pg, 30 pg, 10 pg for **Identifiler** and **Identifiler Plus**)
  - Focus on sample amounts with dropout present to examine stochastic effects including severe imbalance of heterozygous alleles and allele dropout
- Stochastic Threshold: The RFU value of highest surviving false homozygous peak per dye channel


# Heat Map Explanation

Green = full (correct) type  
Yellow = allele dropout  
Red = locus dropout



Results broken down by locus



A single profile slice



A replicate slice



This is an easy way to look at a lot of data at once


n=84 Samples

# Stochastic Threshold

Identifiler: 28 cycles

Standard Injection on 3500:  
7 sec @ 1.2 kV inj

A	Identifiler with independent AT for each dye				D 3 S 1 3 5 8	D 1 3 S 3 1 7	D 1 6 S 5 3 9	D 2 S 1 3 3 8	D 1 9 S 4 3 3	v W A	T P O X	D 1 8 S 5 1	A M E L	D 5 S 8 1 8	F G A	
	D 8 S 1 1 7 9	D 2 1 S 1 1 1	D 7 S 8 2 0	C S F 1 P O												
A																243
50 pg																265
																167
																148
B																
50 pg																
C																
50 pg																
A																
30 pg																
B																
30 pg																
C																
30 pg																



344



409



435



309



# Summary of Thresholds

Both AT and ST values rounded to the nearest 5 RFU value

Expected peak height ratio (PHR) is assuming the possibility of having one peak at the AT and one peak at the ST

**Expected PHR = AT/ST**

<b>Identifiler: 7 sec @ 1.2 kV (28 cycles)</b>				
	<b>AT (RFU)</b>	<b>Highest Surviving Peak (RFU)</b>	<b>ST (RFU)</b>	<b>Expected PHR</b>
<b>Blue</b>	95	344	345	28%
<b>Green</b>	130	435	435	30%
<b>Yellow</b>	140	409	410	34%
<b>Red</b>	120	309	310	39%

<b>Identifiler Plus: 7 sec @ 1.2 kV (28 cycles)</b>				
	<b>AT (RFU)</b>	<b>Highest Surviving Peak (RFU)</b>	<b>ST (RFU)</b>	<b>Expected PHR</b>
<b>Blue</b>	55	288	290	19%
<b>Green</b>	75	383	385	19%
<b>Yellow</b>	105	414	415	25%
<b>Red</b>	120	265	265	45%



# Consumable RFID Tracking Limits

	<b>RFID Hard Stops</b>	<b>Usage Comments From a Research Laboratory Standpoint</b>
<b>Array</b>	None	<ol style="list-style-type: none"> <li>1. Very easy to change between HID and sequencing</li> <li>2. Array from validation was stored at least twice and reinstalled on 3500 during validation</li> </ol>
<b>Buffer</b>	Expiration Date 7 Days on Instrument # Injections	<ol style="list-style-type: none"> <li>1. Can no longer use in-house buffer</li> <li>2. Very easy to change on the instrument (snap-and-go)</li> </ol>
<b>Polymer</b>	Expiration Date # Samples # Injections	<ol style="list-style-type: none"> <li>1. Hard stop with the expiration date has caused us to discard unused polymer we would have otherwise kept on the instrument</li> <li>2. ~50% of total polymer remains in the pouch after “consumption”</li> <li>3. Expiration dates have changed purchasing strategy (smaller batches, based on ongoing project needs)</li> </ol>

# Validation Conclusions

- The 3500 has proven to be reliable, reproducible and robust
  - Out of 498 samples between **Identifiler** and **Identifiler Plus** only 5 required reinjection
- Dye specific analytical thresholds resulted in less allelic and full locus dropout than applying one analytical threshold to all dyes
- Stochastic thresholds are linked to analytical thresholds
  - If the analytical threshold is adjusted, the stochastic threshold should be reevaluated along with expected peak height ratios
    - Requires consideration for overall interpretation workflow which we are still evaluating
- RFID tracking decreases flexibility in our research experience

What is Normalization and  
how does it work?

# Normalization of Data

- Recommended to compare signal between instruments
- Motivation mainly for large laboratories with many instruments
  - Correct for signal variation between instruments
- Can be used with a single instrument
  - Correct for signal variation between single and multiple injections

# Normalization Definitions

- Normalization Target (NT)
  - Requires the use of LIZ 600 v2.0 size standard
  - Average peak heights of 11 peaks within LIZ 600 v2.0 selected for peak height consistency across lots
  - Applied within data collection software prior to running samples

▼ **Advanced Options**  
Following values are not recommended to be changed.

Voltage Tolerance (kVolts):	<input type="text" value="0.7"/>	Voltage # of Steps (nk):	<input type="text" value="20"/>	Voltage Step Interval (sec.):	<input type="text" value="15"/>
First Read Out Time (ms):	<input type="text" value="160"/>	Second Read Out Time (ms):	<input type="text" value="160"/>		
<b>Normalization Target:</b>	<input type="text" value="3200"/>	Normalization Factor Threshold Min:	<input type="text" value="0.3"/>	Normalization Factor Threshold Max:	<input type="text" value="3.0"/>

# Normalization Definitions

- Normalization Factor (NF)
  - Adjustment needed for individual samples to reach the Normalization Target value
  - Full signal adjustment (baseline, peaks, artifacts, etc)
    - Either **increase** or **decrease** signal

## Sample Information

```
Sample File      : Ladder_A01_01.hid
Sample Name     : Ladder
Sample Origin Path : C:\Documents and Settings\ericab\My Documents\Erica\3500 Validation\Run Folders\3500\Normalization\Mixtures\Run
2011-05-12-10-45-50-422\Identifiler\Inj1 2011-05-12-10-48-10-679\Ladder_A01_01.hid
Status Message  : Analyzed
File Source     : Disk media
Re-Injection    : NA
Assay Name      : Identifiler
Assay Version   : v1.0.0
Normalization Factor : 0.995
```


# Normalization Example

Theoretical Normalization Target: 2000 RFU

Without Normalization

With Normalization


LIZ 600 v2.0  
Peak Height  
Average:  
**1021 RFU**




NF=1.959




Signal increases  
almost by a factor of 2




LIZ 600 v2.0  
Peak Height  
Average:  
**3192 RFU**



NF= 0.596



Signal decreases  
by almost half



# Future Work

- Validation of additional kits (Promega)
- More extensive review of the impact of thresholds on interpretation
  - Interaction between analytical and stochastic thresholds alongside peak height ratios
- More extensive review of normalization
  - Do thresholds change when employing normalization?



# Acknowledgments

## *Forensic DNA Team*



John  
Butler



Mike  
Coble



Becky  
Hill



Margaret  
Kline

## *Data Analysis Support*



Dave  
Duewer

## *DNA Biometrics Team*



Pete  
Vallone



Kristen Lewis  
O'Connor

Funding from the **National Institute of Justice (NIJ)**  
through NIST Office of Law Enforcement Standards

Funding from the **FBI Biometrics Center  
of Excellence** 'Forensic DNA Typing as a  
Biometric Tool'

Jeff Sailus-Applied Biosystems

Contact Info:

[erica.butts@nist.gov](mailto:erica.butts@nist.gov)

+1-301-975-5107

